



Dynalab Corp. *Dynalon Labware*

350 Commerce Drive
Rochester, New York 14623 USA

Tel: 585-334-2060 800-828-6595
Fax:585-334-0241
dynalabcorp.com *dynalon.com*

TECHNICAL DATA SHEET

Product Group		POLYSTYRENE DIPPAS®
Material Composition	Base	Approved General Purpose Virgin Polystyrene
	Caps	Polypropylene or Metal (tin plate or aluminium) coated with lacquer.
	Cap Lining	Flowed seal – non toxic plastic (metal caps only)

ALL THE ABOVE MATERIALS ARE FDA APPROVED.

Manufacture

- 1 All polystyrene dippas are manufactured in compliance with product and quality requirements as documented in the BS EN ISO 9001:2000 Quality System.
- 2 Verification of product clean room conditions in accordance with BS EN ISO 14644 'Environmental Cleanliness in enclosed spaces'
- 3 Designed and leak tested in accordance with British Standard BS EN 14254:ANNEX D
- 4 Available in either clear or blue.
- 5 Terminally sterilised by gamma irradiation.

Product Details

Item Code	Description	Capacity (ml)	Colour	Handle Length (mm)	Materials Base/Cap	Sterility
705-0030	Dippa, screw cap	30	Clear	191	PS/PP	IRR
705-B030	Dippa, screw cap	30	Blue	191	PS/PP	IRR
705-0100	Dippa, screw cap	100	Clear	383	PS/ME	IRR
705-B100	Dippa, screw cap	100	Blue	383	PS/ME	IRR
705-0250	Dippa, screw cap	250	Clear	334	PS/ME	IRR
705-B250	Dippa, screw cap	250	Blue	334	PS/ME	IRR



Quality in Life Sciences



IWAKI

