



**MANUFACTURED IN THE UK**

**Dynalab Corp.** *Dynalon Labware*

350 Commerce Drive  
Rochester, New York 14623 USA

Tel: 585-334-2060 800-828-6595  
Fax: 585-334-0241  
dynalabcorp.com *dynalon.com*

**TECHNICAL DATA SHEET**

**Product Group**

**Polystyrene Petri Dishes**

**Material Composition**

Lid & Base

General Purpose Virgin Polystyrene

ALL THE ABOVE MATERIALS ARE FDA APPROVED.

**Manufacture**

- 1 All Polystyrene Petri Dishes are manufactured in compliance with product and quality requirements as documented in the BS EN ISO 9001:2000 Quality System.
- 2 Verification of product clean room conditions in accordance with BS EN ISO 14644 'Environmental Cleanliness in enclosed spaces'
- 3 Petri Dishes are produced and tested in accordance with the BS611 PART 2 standard, which includes stringent dimensional controls.
- 4 Sterilin Petri Dishes are CE marked conforming to the In-vitro diagnostic directive (98/79/EC)

**Sterilisation**

- 1 Also available in a sterile format, sterilised by gamma irradiation (dosage 10 – 30 kGy)
2. Sterility is obtained through irradiation according to ISO 11137 (sterilisation of healthcare products - requirements for validation and routine control - radiation sterilisation) and EN552 (sterilisation of Medical Devices – validation and routine control of sterilisation by irradiation).



**IWAKI**

