



MANUFACTURED IN THE UK

Dynalab Corp. *Dynalon Labware*

350 Commerce Drive
Rochester, New York 14623 USA

Tel: 585-334-2060 800-828-6595

Fax: 585-334-0241

dynalabcorp.com

dynalon.com

TECHNICAL DATA SHEET

Product Group	POLYSTYRENE PIPETTES
Material Composition	Approved General Purpose Virgin Polystyrene

ALL THE ABOVE MATERIALS ARE FDA APPROVED.

Manufacture

- 1 All polystyrene pipettes are manufactured in compliance with product and quality requirements as documented in the BS EN ISO 9001:2000 Quality System.
- 2 Volume ranges from 1ml to 50ml
- 3 Accuracy of $\pm 1\%$ for nominal volume
- 4 Terminally sterilised by gamma irradiation
- 5 Non pyrogenic with an endotoxin level of below 0.25EU/ml.

Product Details

Sterilin single use, serological pipettes are manufactured using virgin grade polystyrene and to stringent quality specifications.

Each size and version of pipette has clear black printed graduations for maximum clarity; descending and ascending graduations to aid pipetting; negative graduations for extra capacity and a colour coding for ease of identification.